

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Mercury (Hg) ICP Standard at 1,000 µg/mL in 2% HNO3 500 mL HPS-100033-1-500**

|                            |                                                                                                                                                            |
|----------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Article Name               | Mercury (Hg) ICP Standard at 1,000 µg/mL in 2% HNO3 500 mL                                                                                                 |
| Biozol Catalog Number      | HPS-100033-1-500                                                                                                                                           |
| Supplier Catalog Number    | 100033-1-500                                                                                                                                               |
| Alternative Catalog Number | HPS-100033-1-500-500                                                                                                                                       |
| Manufacturer               | High Purity Standards                                                                                                                                      |
| Category                   | Biochemikalien                                                                                                                                             |
| Product Description        | Hg 1000 µg/mL in 2% HNO3 18 months expiry date. Traceable to NIST SRM 3133. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited.... |